

Original paper

## Computed tomography independent quantitative determinants of CAD-RADS 4 versus CAD-RADS 3 for calcified coronary lesions

Amal Abdelsattar Sakrana<sup>1,A,B,E</sup>, Heba M. Abou El Atta<sup>2,B,D</sup>, Goda Mohammad<sup>3,B,E</sup>, Dalia Bayoumi<sup>1,B,D</sup>

<sup>1</sup>Department of Diagnostic and Interventional Radiology, Mansoura University Hospital, Egypt

<sup>2</sup>Department of Radiology, Student Hospital, Mansoura University, Egypt

<sup>3</sup>Department of Cardiology, Zagazig University Hospital, Egypt

### Abstract

**Purpose:** An investigation of coronary computed tomography angiography (CCTA)-derived quantitative parameters to determine CAD-RADS 4 versus CAD-RADS 3 of coronary lesions with moderate to severe calcification.

**Material and methods:** The study included 150 coronary lesions proven to have moderate or severe stenosis by invasive coronary angiography and showing moderate to severe calcification in CCTA. Various CCTA-quantitative parameters were correlated to the degree of stenosis (moderately versus severely stenosed lesions). Their sensitivity and specificity to detect severe stenosis (supposed to be corresponding to CAD-RADS 4) were examined at multiple cut-off points.

**Results:** The calcification remodelling index (CRI) was the only statistically significant independent computed tomography angiography-derived predictor of severe stenosis versus moderate stenosis on multivariate regression analysis. The best cut-off value was  $\leq 0.84$ , with 77.78% sensitivity and 86.46% specificity.

**Conclusions:** From all quantitative-derived CCTA parameters,  $\text{CRI} \leq 0.84$  was the predictor with the highest diagnostic performance for severe versus moderate stenosis in moderately to severely calcified coronary lesions. Accordingly, CRI can help to determine CAD-RADS 4 versus CAD-RADS 3.

**Key words:** coronary stenosis, coronary CT angiography, coronary calcification, calcified coronary lesions, calcified plaques, CAD-RADS.

### Introduction

Coronary computed tomography angiography (CCTA) has been well validated as a non-invasive tool for the evaluation of coronary artery disease (CAD) with high sensitivity and negative predictive value [1,2]. However, the accuracy of CCTA to assess coronary luminal stenosis is typically reduced in the case of moderately to severely calcified coronary lesions [3,4].

On the other hand, CCTA is valuable for the characterization of moderately and heavily calcified coronary lesions

to predict the use of rotational atherectomy before percutaneous coronary intervention (PCI) or stent implantation [5].

The CAD Reporting and Data System (CAD-RADS) is a new collaborative standardized reporting effort of various societies. It is an attempt to decrease the variability among computed tomography (CT) readers, enhance the communication between the referring and interpreting physician, and facilitate uniform recommendations for further management. However, it has some limitations [6]. One of the important challenges is quantifying luminal stenosis with the densely calcified lesion, which usually

### Correspondence address:

Amal Abdelsattar Sakrana, Department of Diagnostic and Interventional Radiology, Mansoura University Hospital, Egypt, e-mail: [amlsakrana@yahoo.com](mailto:amlsakrana@yahoo.com)

### Authors' contribution:

A Study design · B Data collection · C Statistical analysis · D Data interpretation · E Manuscript preparation · F Literature search · G Funds collection

causes overestimation of stenosis due to the blooming and beam-hardening artifacts [7].

Therefore, we assessed a group of quantitative CCTA parameters of moderately and severely calcified coronary lesions diagnosed as moderately (50-69% diameter stenosis) and severely stenosed lesions (70-99% diameter stenosis) by invasive coronary angiography (ICA). We investigated the cut-off values of these quantitative parameters with the highest accuracy to differentiate between moderate and severe stenosis, which will help to designate CAD-RADS 3 and 4 in the future.

## Material and methods

### Patients

Eighty-four patients with 150 calcified coronary lesions, who underwent CCTA and ICA with one-month intervals, were included retrospectively in our study. The inclusion criteria of the study lesions were [1] moderately (calcification arc angle [CAA] = 90-179°) or severely (CAA = 180-360°) calcified coronary lesions on CCTA, and [2] ICA-diagnosed cases as moderately stenosed (50-69%) or severely stenosed (70-99%) lesions. We excluded the following lesions from our study: ICA-diagnosed lesions as mildly stenosed lesions or occluded lesions, vessel segments of diameter < 1.5 mm, tortuous vessel segments, distal vessel segments, vessel segments distal to occlusion, stenotic lesions due to mixed plaques. Also, we excluded cases with prior coronary arteries bypass surgery (CABG) and patients with renal dysfunction glomerular filtration rate (GFR) < 45 ml/min. Each patient gave written informed consent. Institutional Ethics Committee approval was obtained for the study.

### Method

CT Image acquisition CT was performed on a multidetector CT scanner (Philips Ingenuity Core 128, Philips Healthcare, Netherlands). Patients with heart rate > 65 bpm were given oral  $\beta$ -blockers (25-75 mg metoprolol) before the examination. Before CCTA, coronary artery calcium scoring (CAC) was performed with prospective electrocardiography (ECG)-triggered 3-mm-thickness axial slices in the diastolic phase without overlapping or gaps on the cardiac region using a medium-sharp convolution Kernel. All patients received an intravenous bolus dose of 80 to 120 cc of contrast agent (Iodixanol 350 mg iodine/ml) according to each patient's weight and scan time. The contrast flow rate was 5 ml/s. Contrast injection was followed by a 40 cc saline bolus. A bolus tracking technique was used to trigger scanning with the proper time delay (100 Hounsfield unit [HU] threshold in ascending aorta ROI, and 4 s after the peak). Prospectively, ECG-triggered CCTA was performed with detector collimation of  $64 \times 0.625$  mm with a z-flying focal spot, rotation time

of 300 ms, temporal resolution of 150 ms, and reconstructed slice interval of 0.5 mm. The tube current was modified automatically with the patient-specific size and attenuation of the body region via the CARE dose system. The ECG-dependant dose modulation technique with full current dose was applied in the R-R interval of 40-70%. The tube voltage was adjusted to 100 kV for patients having body mass index (BMI) < 26 kg/m<sup>2</sup> and 120 kV for those having BMI > 26 kg/m<sup>2</sup>.

### Computed tomography image reconstruction

Images were post-processed on a dedicated workstation (IntelliSpace Portal V5, Philips Healthcare). We calculated the segmental coronary calcium score (SCCS) and the segmental calcium volume (SCV) on CAC scoring. According to the Agatston method [8], calcification is considered if its area is > 1 mm<sup>2</sup> and the peak CT number (CTN) is > 130 HU. We determined the SCCS by multiplying the calcification area by the peak CTN for each included calcified coronary lesion. The SCV also was calculated automatically for each included calcified lesion.

From CCTA, axial images were reconstructed with 0.6 mm slice thickness and an increment of 0.5 mm using a medium-sharp kernel. The coronary vessels were then evaluated on curved planar reformations (CPR), interactive oblique multiplane reformations (MPR), and cross-sectional planes.


The calcified lesion length (LL) was measured on each lesion's best long-axis plane from its proximal to distal end. On coronary vessel cross-sectional images we calculated the CT-diameter stenosis percentage (CT-DS %), calcification remodelling index (CRI), and calcification arc angle, as shown in (Figure 1), using the following formula:

$$\text{CT-DS\%} = \frac{\text{normal proximal/distal reference diameter} - \text{vessel diameter at the most severely stenosed point}}{\text{normal proximal/distal reference diameter}} \times 100$$

The normal vessel reference was selected manually. CRI was calculated as the luminal cross-sectional area ratio at the most severely calcified site (including the calcium) to that of the proximal reference. CAA was measured as the central angle subtended by the calcification arc in the cross-sectional plane.

### Invasive coronary angiography

ICA was performed within a 30-day interval from CCTA via femoral access using conventional techniques on our fluoroscopy unit (Axiom, Siemens Medical Systems). The main coronary vessels were displayed in the multiple standard projections. An expert interventional cardiologist blindly and independently estimated the stenosis extent semi-automatically. A coronary lesion of  $\geq 70\%$  luminal diameter narrowing was defined as severe stenosis, while moderate stenosis was defined for the lesions of 50-69% diameter reduction.



**Figure 1.** Coronary computed tomography angiography-derived measurements of a severely stenosed calcified left anterior descending artery (LAD) lesion. Axial oblique maximal intensity projection (MIP) (A). Axial oblique MIP with coloured plaque characterization (B). CT-calcium scoring (C). LAD-cross sectional multiplanar reformates (MPR) (D-G). LAD-curved MPR (H). Invasive coronary angiography (ICA) cranial left anterior oblique (LAO) view. ICA shows proximal LAD severely stenosed lesion appears as a heavily calcified lesion on CCTA (F). Image illustrates calcification arc angle (CAA) measurements (E and G). Images demonstrate the calculation of calcification remodelling index (CRI). SCS – segmental calcium score, AU – Agatston units, SCV – segmental calcium volume

### Statistical analysis

Data were analysed using Statistical Package for the Social Sciences software version 26 (SPSS, Inc., Chicago, IL, USA). The Shapiro-Wilk test was used to assess normality of distribution of data. Data were presented as median and interquartile range (IQR). The Mann-Whitney test

was used to compare the 2 groups (moderate and severe stenosis) with quantitative nonparametric data. Univariate and multivariate logistic regression was used to detect predictors of stenosis extent. The sensitivity and specificity of CT-DS, CRI, SCS, SCV, CL, and CAA to differentiate between moderate and severe stenosis groups were examined at different cut-off points using receiver operat-

ing characteristic (ROC) curve analysis to determine the best cut-off point as well as the diagnostic power of each test. *P*-values less than 0.05 were considered statistically significant.

## Results

A total of 84 patients with 150 calcified coronary lesions were included in the present study; 99 (66%) were ICA-diagnosed as moderately stenosed lesions, and 51 (34%) were severely stenosed. The study population included 58 men (69%), mean age:  $70.6 \pm 8.3$  years, mean BMI:  $28.5 \pm 5.4$  kg/m<sup>2</sup>. Table 1 summarizes the detailed clinical characteristics of the study populations. Most of the included coronary lesions involved the left anterior descending (LAD) artery, and the proximal vessel segments were involved more than the middle vessel segments.

As shown in Table 2, and Figure 2, severe coronary stenotic lesions ( $\geq 70\%$  luminal stenosis by ICA) displayed significantly larger CT-DS%, SCS, SCV, and CL as well as smaller CRI. CAA was larger with severely stenotic lesions, but not statistically significantly.

**Table 1.** Study demographic and clinical characteristics

Characteristic	Values
Number of patients	84
Number of lesions (%)	
Total	150 (100)
ICA-diagnosed moderately stenosed lesions	99 (66)
ICA-diagnosed severely stenosed lesions	51 (34)
Age, years	$70.6 \pm 8.3$
Gender (male), <i>n</i> (%)	58 (69)
BMI (kg/m <sup>2</sup> )	$28.5 \pm 5.4$
Family history of coronary artery disease, <i>n</i> (%)	11 (9.2)
Smoking, <i>n</i> (%)	54 (64.3)
Diabetes mellitus, <i>n</i> (%)	40 (47.6)
Hypertension, <i>n</i> (%)	39 (46.4)

**Table 2.** Coronary computed tomography angiography – parameters of severe versus moderate stenosis

	Invasive coronary angiography – stenosis extent			<i>p</i> -value
	Moderate stenosis, <i>n</i> = 99	Severe stenosis, <i>n</i> = 51	Total, <i>N</i> = 150	
CT-DS (%)	78.0 (65.3-89.5)	91.5 (75.2-96.2)	81.0 (67.9-93.2)	0.04*
CRI	0.94 (0.86-1.08)	0.68 (0.57-0.83)	0.88 (0.79-1.00)	< 0.001*
SCS (AU)	229 (118-415)	246 (121-458)	241 (121-420)	0.01*
SCV (mm <sup>3</sup> )	199.1 (92.7-325.1)	210.3 (94.9-376.4)	206.5 (94.5-327.6)	0.03*
CL (mm)	14.6 (10.2-23.9)	15.3 (10.7-25.6)	15.1 (10.5-24.2)	0.03*
CAA (°)	220 (157-268)	245 (176-320)	234 (169-286)	0.098

Data expressed as median (IQR). \*Significance < 0.05. Test used: Mann-Whitney

CT-DS – diameter stenosis % by CT, CRI – calcification remodelling index, SCS – segmental calcium score, AU – Agatston units, SCV – segmental calcium volume, CL – calcification length, CAA – calcification arc angle

Table 3 shows the results of univariate and multivariate regression analysis of CCTA-derived quantitative data in both ICA-based stenosis degrees. As regards the univariate analysis, CRI was significantly lower in severely stenosed calcified coronary lesions, while the remaining CT-derived quantitative parameters were significantly higher in severely stenosed coronary lesions. In multivariate regression analysis, CRI was the only independent determinate of ICA-diagnosed stenosis degree (*p* < 0.001).


ROC analysis and the diagnostic performance of CTA-derived parameters to predict ICA-diagnosed, severely stenosed, calcified coronary lesions are displayed in Table 4 and Figure 3. The highest diagnostic performance appertained to CRI with an area under the curve and 95% confidence interval (AUC [95% CI]) of 0.857 (0.79-0.91).  $CRI \leq 0.84$  had the highest sensitivity (77.78%) and specificity (86.46%).

## Discussion

Many previous studies investigated the diagnostic accuracy of CCTA in the evaluation of coronary stenosis extent caused by calcified plaques [3,4,9-11]. Blooming artifacts have always been the supposed main hindering factor of accurate assessment of the vessel stenosis [12].

The accurate prediction of severe coronary stenosis in moderately to severely calcified coronary lesions is still challenging. In their study, Qi *et al.* concluded that CCTA showed higher diagnostic accuracy for coronary artery stenosis when the ratio of calcium plaque volume to the vessel circumference (RVTC) was  $\leq 50\%$ , while it displayed significantly decreased accuracy with larger RVTC [4]. Krub *et al.* stated in their study that a mean lumen diameter of  $\leq 2.8$  mm along with a total calcium arc of  $\geq 47^\circ$  predict the lumen underestimation by CCTA [11]. We included in our study moderately and severely calcified coronary lesions having CAA  $\geq 90^\circ$ .

Few studies have investigated quantitative CCTA parameters that could predict the extent of stenosis with the calcified coronary lesions [13]. We investigated different quantitative CCTA parameters measured in ICA-



**Figure 2.** Box plot of coronary computed tomography angiography (CCTA)-derived parameters in moderate stenosis versus severe stenosis. \*Significance < 0.05

**Table 3.** Univariate and multivariate analysis of coronary computed tomography angiography (CCTA) parameters in severe coronary stenosis versus moderate stenosis

	All, N = 150	Moderate stenosis, n = 99	Severe stenosis, n = 51	Univariate analysis, p-value	Multivariate analysis, p-value
CT-DS (%)	81.0 (67.9-93.2)	78.0 (65.3-89.5)	91.5 (75.2-96.2)	0.039*	0.496
CRI	0.88 (0.79-1.00)	0.94 (0.86-1.08)	0.68 (0.57-0.83)	< 0.001*	< 0.001*
SCS (AU)	241 (121-420)	229 (118-415)	246 (121-458)	0.029*	0.719
SCV (mm <sup>3</sup> )	206.5 (94.5-327.6)	199.1 (92.7-325.1)	210.3 (94.9-376.4)	0.019*	0.254
CL (mm)	15.1 (10.5-24.2)	14.6 (10.2-23.9)	15.3 (10.7-25.6)	0.048*	0.380
CAA (°)	234 (169-286)	220 (157-268)	245 (176-320)	0.048*	0.710


\*Significance < 0.05.

**Table 4.** Performance of coronary computed tomography angiography (CCTA) parameters in discrimination of severe coronary stenosis

	AUC (95% CI)	p-value	Cut off point	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)	Accuracy (%)
CT-DS (%)	0.6 (0.5-0.7)	0.04*	78.30	58.8	50.5	38.0	70.4	53.3
CRI	0.86 (0.79-0.91)	< 0.001*	0.84	77.8	86.5	76.4	87.4	64.2
SCS (AU)	0.62 (0.53-0.72)	0.01*	238.00	62.7	54.5	41.6	74.0	57.3
SCV (mm <sup>3</sup> )	0.61 (0.51-0.70)	0.03*	196.50	58.8	49.5	37.5	70.0	52.7
CL (mm)	0.61 (0.51-0.70)	0.03*	14.65	62.7	50.5	39.5	72.5	54.7
CAA0 (°)	0.58 (0.48-0.68)	0.09	220.50	62.7	50.5	39.5	72.5	54.7

\*Significance < 0.05

AUC – area under a curve, PPV – positive predictive value, NPV – negative predictive value



**Figure 3.** Reciprocal operative curve (ROC) representing the specificity and sensitivity of coronary computed tomography angiography (CCTA)-derived parameters to identify severe coronary stenosis

diagnosed moderately (50-69%) versus severely (70-99%) stenosed coronary lesions.

The CAD-RADS terminology was first introduced in 2016 by Cury *et al.* as the first scoring system facilitating standardized CAD classification with the potential to improve contact between referring physicians and CCTA imagers [14]. Since then, the score's widespread endorsement has not been established and is still in an early phase of implementation into routine daily clinical practice [15]. CAD-RADS requires CCTA-derived information about the stenosis extent. In an attempt to get quantitative predictors of the stenosis extent, i.e. moderate stenosis (corresponding to CAD-RADS 3) versus severe stenosis (corresponding to CAD-RADS 4) in moderately and severely calcified coronary lesions, we analysed some CCTA-derived quantitative parameters.

In the current study, ICA-diagnosed severe coronary stenosis caused by calcified plaques was significantly associated with smaller CRI than moderate coronary stenosis, while CT-DS%, SCS, SCV, and CL were significantly larger in severe stenosis. Among all CT-derived quantitative parameters, CRI was the only statistically significant independent

predictor of severe stenosis. CRI ≤ 0.84 was the best cut-off value predicting severe coronary stenosis, with 77.78% sensitivity, 86.46% specificity, and AUC (95% CI) 0.857 (0.79-0.91). Yu *et al.* reported similar results, but they reported a larger cut-off value (≤ 0.94). We thought that this difference could be due to the inclusion of mildly stenosed lesions (< 50%) in their study, in contrast to our research, including only more than moderately stenosed lesions [13].

Previously a reasonable explanation was reported of inversely correlated stenosis extent with CRI: it was suggested that calcified coronary lesions with negative remodelling (small CRI) could have intimal locations with more encroachment on the lumen with a subsequently greater degree of stenosis [5,13].

The study limitation was that we did not perform the fractional flow reserve (FFR) for most of the lesions to determine the lesions' haemodynamic significance. However, the study strength was that the inclusion criteria were remarkably selective, allowing optimization of the study sample's homogeneity, reduction of confounding, and increased probability of getting a real association between the study variables.

As a clinical implementation of our study, CRI may help to estimate severe versus moderate stenosis in coronary lesions with moderate to severe calcification at the current stage. Therefore, it can help to assign calcified coronary lesions CAD-RADS 4 versus CAD-RADS 3.

## Conclusions

CRI  $\leq 0.84$  was an independent predictor of severe coronary stenosis versus moderate stenosis in moderately to se-

verely calcified lesions, with 77.78% sensitivity and 86.46% specificity, and hence it can determine CAD-RADS 4 versus CAD-RADS 3.

## Conflict of interest

The authors report no conflict of interest.

## References

1. Hamon M, Morello R, Riddell JW, et al. Coronary arteries: diagnostic performance of 16-versus 64-section spiral CT compared with invasive coronary angiography – meta-analysis. *Radiology* 2007; 245: 720-731.
2. Li Y, Zhang J, Lu Z, et al. Discrepant findings of computed tomography quantification of minimal lumen area of coronary artery stenosis: correlation with intravascular ultrasound. *Eur J Radiol* 2012; 81: 3270-3275.
3. Vavere AL, Arbab-Zadeh A, Rochitte CE, et al. Coronary artery stenoses: accuracy of 64-detector row CT angiography in segments with mild, moderate, or severe calcification – a subanalysis of the CORE-64 trial. *Radiology* 2011; 261: 100-108.
4. Qi L, Tang L J, Xu Y, et al. The diagnostic performance of coronary CT angiography for the assessment of coronary stenosis in calcified plaque. *PLoS One* 2016; 11: e0154852.
5. Yu M, Li Y, Li W, et al. Calcification remodeling index characterized by cardiac CT as a novel parameter to predict the use of rotational atherectomy for coronary intervention of lesions with moderate to severe calcification. *Korean J Radiol* 2017; 18: 753-762.
6. Ramanathan S, Al Heidous M, Alkuwari M. Coronary artery disease-reporting and data system (CAD-RADS): strengths and limitations. *Clin Radiol* 2019; 74: 411-417.
7. Zhang S, Levin DC, Halpern EJ, et al. Accuracy of MDCT in assessing the degree of stenosis caused by calcified coronary artery plaques. *Am J Roentgenol* 2008; 191: 1676-1683.
8. Agatston AS, Janowitz WR, Hildner FJ, et al. Quantification of coronary artery calcium using ultrafast computed tomography. *J Am Coll Cardiol* 1990; 15: 827-832.
9. Cerci R, Vavere AL, Miller JM, et al. Patterns of coronary arterial lesion calcification by a novel, cross-sectional CT angiographic assessment. *Int J Cardiovasc Imaging* 2013; 29: 1619-1627.
10. Palumbo AA, Maffei E, Martini C, et al. Coronary calcium score as gatekeeper for 64-slice computed tomography coronary angiography in patients with chest pain: per-segment and per-patient analysis. *Eur Radiol* 2009; 19: 2127-2135.
11. Kruk M, Noll D, Achenbach S, et al. Impact of coronary artery calcium characteristics on accuracy of CT angiography. *JACC Cardiovasc Imaging* 2014; 7: 49-58.
12. Pundziute G, Schuijff JD, Jukema JW, et al. Impact of coronary calcium score on diagnostic accuracy of multislice computed tomography coronary angiography for detection of coronary artery disease. *J Nucl Cardiol* 2007; 14: 36-43.
13. Yu M, Li Y, Li W, et al. Calcification remodeling index assessed by cardiac CT predicts severe coronary stenosis in lesions with moderate to severe calcification. *J Cardiovasc Comput Tomogr* 2018; 12: 42-49.
14. Cury RC, Abbara S, Achenbach S, et al. CAD-RADSTM coronary artery disease – reporting and data system. An expert consensus document of the Society of Cardiovascular Computed Tomography (SCCT), the American College of Radiology (ACR) and the North American Society for Cardiovascular Imaging (NASCI). Endorsed by the American College of Cardiology. *J Cardiovasc Comput Tomogr* 2016; 10: 269-281.
15. Foldyna B, Szilveszter B, Scholtz JE, et al. CAD-RADS – a new clinical decision support tool for coronary computed tomography angiography. *Eur Radiol* 2018; 28: 1365-1372.